



Test Report

Date : 2023-04-11
No. : HC23031178

Page 1 of 2

Applicant (Code:01325334) : Art Beauty Group Company Limited
Unit 2305 Apec Plaza
43 Hoi Yuen Road
Kwun Tong NT HK

Description of Sample(s) : One submitted sample said to be Disposable Latex Medical Examination Gloves.

Sample(s) Received Condition(s): In intact original package under ambient temperature

Date Sample(s) Received : 2023-03-28

Date Tested : 2023-03-29 to 2023-04-05

Investigation Requested :

1. Total bacterial count
2. Total fungal count
3. Coliforms
4. Hemolytic streptococcus
5. *Pseudomonas aeruginosa*
6. *Staphylococcus aureus*

Conclusion(s) : The test results of the submitted sample **complied** with the limits of GB15979-2002 Hygienic Standard for Disposable Sanitary Products.

LAU Yuk Kuen, Joey
Authorized Signatory
Chemical and Food Department
For and on behalf of

The Hong Kong Standards and Testing Centre Ltd.



The Hong Kong Standards and Testing Centre Limited

10 Dai Wang Street, Tai Po Industrial Estate, Tai Po, N.T., Hong Kong

Tel: +852 2666 1888 Fax: +852 2664 4353 Email: hkstc@stc.group Website: www.stc.group

This report shall not be reproduced unless with prior written approval from The Hong Kong Standards and Testing Centre Limited.

For Conditions of Issuance of this test report, please refer to the overleaf and Website.



Test Report

Date : 2023-04-11
No. : HC23031178

Page 2 of 2

Method(s) Used:

GB15979-2002 Hygienic Standard for Disposable Sanitary Products, Appendix B.

Test Result(s):

Test Item(s)	Disposable Latex Medical Examination Gloves	Microbial Limit (GB 15979-2002) *
1. Total bacterial count	<10 CFU/g	≤200 CFU/g
2. Total fungal count	<20 CFU/g	≤100 CFU/g
3. Coliforms	Not detected	Not detected
4. Hemolytic streptococcus	Not detected	Not detected
5. <i>Pseudomonas aeruginosa</i>	Not detected	Not detected
6. <i>Staphylococcus aureus</i>	Not detected	Not detected

Notes : - < denotes less than
- CFU/g denotes colony forming unit per gram
- ≤ denotes less than or equal to
- *These limits are only applicable to products that are in intact original package. Products that are already in use or have their package broken are not included.
- Decision Rule: for numerical result with upper limit, compliance is deemed to occur if the measured result is under the limit, even extended upwards by the expanded uncertainty with 95% coverage probability.

***** End of Test Report *****

The Hong Kong Standards and Testing Centre Limited

10 Dai Wang Street, Tai Po Industrial Estate, Tai Po, N.T., Hong Kong

Tel: +852 2666 1888 Fax: +852 2664 4353 Email: hkstc@stc.group Website: www.stc.group

This report shall not be reproduced unless with prior written approval from The Hong Kong Standards and Testing Centre Limited.

For Conditions of Issuance of this test report, please refer to the overleaf and Website.