



香港標準及檢定中心
Hong Kong Standards and Testing Centre

“STC tested” Mark


License Number: STM3911

Date of Issue: 2020-08-04

Issued To: Art Beauty Group Company Limited
Unit 2305, Apec Plaza,
49 Hoi Yuen Road
Kwun Tong, Kowloon

The product listed below is eligible to bear the “STC tested” Mark (優質「正」印)

Issued By: Richard Fung
Chief Executive



For and on behalf of
Hong Kong Standards and
Testing Centre



Product Category

Brand / Product Name

Household Product

Surgical Face Mask

Applicable standard(s) / requirements

1. The submitted sample Surgical Face Mask is able to filter out $\geq 98\%$ biological aerosol of *Staphylococcus aureus* with average particle size: 2.7 μ m (Bacterial Filtration Efficiency, BFE).
2. The test results of the submitted sample complied with the limit of BS EN 14683:2019 Differential Pressure <40 Pa/cm²
3. The test results of the submitted sample complied with the limit of BS EN 14683:2019 Microbial cleanliness (Bioburden)
4. The test results of the submitted sample complied with the limits of GB15979-2002 Hygienic Standard for Disposable Sanitary Products.

Report Ref. No. HC20061073(S)

The use of this License is governed by the “STC tested” Mark Regulations and any other requirement accompanying this License when issued, as amended from time to time.
This License is the property of Hong Kong Standards and Testing Centre and must be returned within one month from occurrence of user’s breach of the aforesaid rules.

Original Certification: 2020-08-04

Expiry: 2021-08-04

Cat. [V]

香港新界大埔工業邨大宏街10號 10 Dai Wang Street, Taipo Industrial Estate, N.T., Hong Kong

Tel: (852) 2666 1887 Fax: (852) 2664 4353 Website: www.stc-group.org Email: stcmark@hkstc.org



香港標準及檢定中心
Hong Kong Standards and Testing Centre

“STC tested” Mark

License Number: STM4041

Date of Issue: 2020-11-05

Issued To: Art Beauty Group Company Limited
Unit 2305 Apec Plaza
49 Hoi Yuen Road
Kwun Tong Kln HK

The product listed below is eligible to bear the “STC tested” Mark (優質「正」印)

Issued By: Richard Fung
Chief Executive



For and on behalf of
Hong Kong Standards and
Testing Centre



Product Category

Brand / Product Name

Household Product

Surgical Face Mask (Child)

Applicable standard(s) / requirements

1. The submitted sample Surgical Face Mask (Child) is able to filter out $\geq 98\%$ biological aerosol of *Staphylococcus aureus* with average particle size: $2.7\mu\text{m}$ (Bacterial Filtration Efficiency, BFE).
2. The test results of the submitted sample complied with the limit of BS EN 14683:2019 Differential Pressure $< 40 \text{ Pa/cm}^2$
3. The test results of the submitted sample complied with the limit of BS EN 14683:2019 Microbial cleanliness (Bioburden)
4. The test results of the submitted sample complied with the limits of GB15979-2002 Hygienic Standard for Disposable Sanitary Products.

Report Ref. No. HC20061073, HC20090592.

The use of this License is governed by the “STC tested” Mark Regulations and any other requirement accompanying this License when issued, as amended from time to time. This License is the property of Hong Kong Standards and Testing Centre and must be returned within one month from occurrence of user’s breach of the aforesaid rules.

Original Certification: 2020-11-05

Expiry: 2021-11-04

Cat. [V]

香港新界大埔工業邨大宏街10號 10 Dai Wang Street, Taipo Industrial Estate, N.T., Hong Kong

Tel: (852) 2666 1887 Fax: (852) 2664 4353 Website: www.stc-group.org Email: stcmark@hkstc.org