



Test Report

Date: 2020-08-19
No.: HC20080670

Page 1 of 2

Applicant(Code:01325334) : Art Beauty Group Company Limited
Unit 2305 Apec Plaza
49 Hoi Yuen Road Kwun Tong Kln

Description of Sample(s) : One submitted sample said to be Surgical Face Mask.
Country of Origin : China
Sample Receiving Condition: In intact original packing under ambient temperature

Date Sample(s) Received : 2020-07-21

Date Tested : 2020-07-22 to 2020-08-12

Investigation Requested : Bacterial Filtration Efficiency 細菌過濾效率 (BFE) %
– *Staphylococcus aureus* 金黃色葡萄球菌 (ATCC 6538)

Method(s) Used : ASTM F2100-2019 9.1 & ASTM F2101-2019

LAU Yuk Kuen, Joey
Authorized Signatory
Chemical and Food Department
For and on behalf of



The Hong Kong Standards and Testing Centre Ltd.

Note: When a statement of conformity to a specification or standard is provided, the ILAC-G8 Guidance document (and/or IEC Guide 115 in the electrotechnical sector) will be adopted as a decision rule for the determination of conformity unless it is inherent in the requested specification or standard, or otherwise specified in the Report.

The Hong Kong Standards and Testing Centre Limited

10 Dai Wang Street, Tai Po Industrial Estate, Tai Po, N.T., Hong Kong

Tel: +852 2666 1888 Fax: +852 2664 4353 Email: hkstc@stc.group Website: www.stc.group

This report shall not be reproduced unless with prior written approval from The Hong Kong Standards and Testing Centre Limited.

For Conditions of Issuance of this test report, please refer to the overleaf and Website.



Test Report

Date : 2020-08-19
No. : HC20080670

Page 2 of 2

Test Result(s):

Bacterial Filtration Efficiency 細菌過濾效率 (BFE) %
– *Staphylococcus aureus* 金黃色葡萄球菌 (ATCC 6538)

Specimen(s)	Surgical Face Mask
1	99.3%
2	99.1%
3	99.2%
4	99.1%
5	99.7%

Notes : - Control average 對照組的平均菌落數 : 2021 CFU

- Mean particle size 平均粒徑 : 2.8 μm

- Testing side 測試面 : Outside of specimen 外側

- Testing area 測試面積 : 47 cm^2

***** End of Test Report *****

The Hong Kong Standards and Testing Centre Limited

10 Dai Wang Street, Tai Po Industrial Estate, Tai Po, N.T., Hong Kong

Tel: +852 2666 1888 Fax: +852 2664 4353 Email: hkstc@stc.group Website: www.stc.group

This report shall not be reproduced unless with prior written approval from The Hong Kong Standards and Testing Centre Limited.

For Conditions of Issuance of this test report, please refer to the overleaf and Website.