



Test Report

Date: 2020-11-19
No.: HC20110356

Page 1 of 2

Applicant(Code:01325334) : Art Beauty Group Company Limited
Unit 2305 Apec Plaza
43 Hoi Yuen Road
Kwun Tong Kln HK

Description of Sample(s) : One submitted sample said to be Surgical Face Mask.
Country of Origin: China

Sample Receiving Condition: In plastic bag under ambient temperature

Date Sample(s) Received : 2020-11-11

Date Tested : 2020-11-12 to 2020-11-19

Investigation Requested : GB 15979 Hygienic Standard for Disposable Sanitary Products
- Total bacterial count
- Total fungal count
- Coliforms
- Hemolytic streptococcus
- *Pseudomonas aeruginosa*
- *Staphylococcus aureus*

Conclusion(s) : The test results of the submitted sample **complied** with the limits of GB15979-2002 Hygienic Standard for Disposable Sanitary Products.

LAU Yuk Kuen, Joey
Authorized Signatory
Chemical and Food Department
For and on behalf of

The Hong Kong Standards and Testing Centre Ltd.



The Hong Kong Standards and Testing Centre Limited

10 Dai Wang Street, Tai Po Industrial Estate, Tai Po, N.T., Hong Kong

Tel: +852 2666 1888 Fax: +852 2664 4353 Email: hkstc@stc.group Website: www.stc.group

This report shall not be reproduced unless with prior written approval from The Hong Kong Standards and Testing Centre Limited.

For Conditions of Issuance of this test report, please refer to the overleaf and Website.



Test Report

Date : 2020-11-19

Page 2 of 2

No. : HC20110356

Method(s) Used:

GB 15979-2002 Appendix B

Test Result(s):

GB 15979 Hygienic Standard for Disposable Sanitary Products

Test Item(s)	Surgical Face Mask	Microbial Limit (GB 15979-2002) *
1. Total bacterial count	<20 CFU/g	≤200 CFU/g
2. Total fungal count	<20 CFU/g	≤100 CFU/g
3. Coliforms	Not detected	Not detected
4. Hemolytic streptococcus	Not detected	Not detected
5. <i>Pseudomonas aeruginosa</i>	Not detected	Not detected
6. <i>Staphylococcus aureus</i>	Not detected	Not detected

Notes: - CFU/g denotes colony forming unit per gram

- < denotes less than

- ≤ denotes less than or equal to

- *These limits are only applicable to products that are in intact original package. Products that are already in use or have their package broken are not included.

***** End of Test Report *****

The Hong Kong Standards and Testing Centre Limited

10 Dai Wang Street, Tai Po Industrial Estate, Tai Po, N.T., Hong Kong

Tel: +852 2666 1888 Fax: +852 2664 4353 Email: hkstc@stc.group Website: www.stc.group

This report shall not be reproduced unless with prior written approval from The Hong Kong Standards and Testing Centre Limited.

For Conditions of Issuance of this test report, please refer to the overleaf and Website.